

NATtrol™ Influenza/RSV Verification Panel

Catalog Number: NATFRVP-C

PRODUCT DESCRIPTION:

NATtrol™ Influenza/RSV Verification Panel (P/N NATFRVP-C)* is formulated with purified, intact virus particles that have been chemically modified to render them non-infectious and refrigerator stable. NATFRVP-C panel contains 21 x 0.5 mL vials each containing viral NATtrol™ targets listed in the Expected Results table. The panel contains 3 tubes each of the Influenza and RSV positive members and 6 tubes of Coxsackievirus A9 negative member. The panel members are supplied in a purified protein matrix that mimics the composition of a true clinical specimen.

*Pat.: <http://www.zeptometrix.com/patent-information/>

INTENDED USE:

- NATtrol™ Influenza/RSV Verification Panel is designed to evaluate the performance of nucleic acid tests for determination of the presence of viral nucleic acids. NATFRVP-C can also be used for verification of clinical assays, development of diagnostic tests and training of laboratory personnel.
- NATFRVP-C contains intact organisms and should be run in a manner similar to that used for clinical specimens.




ETIOLOGIC STATUS/BIOHAZARD TESTING:

- NATtrol™ inactivation was carried out on each stock used to formulate panel members. The inactivation was verified by the absence of viral growth in validated tissue culture based infectivity assays.
- Purified protein matrix used in the manufacture of this product is treated with 0.09% sodium azide. It was manufactured from materials that have been tested and found non-reactive at the donor level for HIV-1/HIV-2 Antibody, HBsAg and HCV Antibody by FDA licensed donor screening test methods. All materials are also tested for HIV-1 and HCV by FDA approved Nucleic Acid Test (NAT) methods. Heat inactivated bovine based source materials used in the manufacture of this product meet applicable USDA requirements for abattoir sourced animals, traceability and country of origin. The materials were collected at USDA licensed establishments or legally imported from countries recognized by the USDA as negligible or controlled for risk for Bovine Spongiform Encephalopathy (BSE) and other exotic disease agents. Donor animals were inspected ante and post mortem at the abattoir as required by the USDA.

PRECAUTIONS:

- Although NATFRVP-C contains inactivated organisms, it should be handled as if potentially infectious.
- Use Universal Precautions when handling this product.
- To avoid cross-contamination, use separate pipette tips for all reagents.

This product was manufactured in a facility which has a Quality Management System that is ISO 13485 certified.

REF	Catalog Number		Temperature Limitation
LOT	Lot Number		Expiration Date
RUO	For Research Use Only		Biological Risk

RECOMMENDED STORAGE:

- NATtrol™ Influenza/RSV Verification Panel should be stored at 2-8°C.

INSTRUCTIONS FOR USE:

- Shake NATtrol™ sample vigorously for 5 seconds.
- Follow the assay manufacturer's instructions for use to process the sample as a clinical sample.

EXPECTED RESULTS:

- Qualitative results are shown in Table 1 below.
- Each laboratory must evaluate the controls and establish their own acceptance criteria.
- The table shown below is for informational purposes only.

Table 1:

Panel Member	Strain	Expected Result
Influenza A H1	A/New Caledonia/20/99	Flu A is positive.
Influenza A H3	A/Brisbane/10/07	Flu A is positive.
Influenza B	B/Florida/02/06	Flu B is positive.
Respiratory Syncytial Virus A	NA	RSV is positive.
Respiratory Syncytial Virus B	CH93(18)-18	RSV is positive.
Coxsackievirus Type A9	NA	Flu A is negative. Flu B is negative. RSV is negative.

DO NOT USE IN HUMANS. FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

These products are intended for research, product development, quality assurance or manufacturing use. These products are NOT intended for use in the manufacture or processing of injectable products subject to licensure under section 351 of the Public Health Service Act or for any other product intended for administration to humans.

ZeptoMetrix Corporation
878 Main Street
Buffalo, NY 14202
800-274-5487
www.zeptometrix.com