HIV Seroconversion Panel Donor No. 64954
Part Number: HIV9011

Donor Demographics: Donor source plasma was collected from a U.S. licensed blood collection facility. Donor was un-treated and asymptomatic throughout the collection period.

PANEL MEMBER	DATE OF DRAW	**CHIRON** HIV-1 bDNA TEST DATE: 8/20/1998	**ROCHE** COBAS PCR Copies/mL TEST DATE: UNK	**ABBOTT** HIV-1 p24 Ag TEST DATE: 8/27/1998	**ABBOTT** HIV-1 p24 Ag NT% TEST DATE: 9/3/1998	**COULTER** HIV-1 p24 Ag TEST DATE: 8/26/1998	**ABBOTT** AxSYM HIV AB ½ EIA TEST DATE: 9/16/1998	**ABBOTT** HIV AB ½ EIA rDNA TEST DATE: 7/21/1998	**ABBOTT** MUREX HIV Ab ½ + O 3RD Generation TEST DATE: UNK
9011-01	09/07/97	55	<20	0.31	NEGATIVE	0.56	0.29	0.170	0.423
9011-02	09/11/97	55	<20	0.28	NEGATIVE	0.48	0.31	0.256	0.446
9011-03	09/16/97	410	<20	0.26	NEGATIVE	0.48	0.30	0.256	0.505
9011-04	09/18/97	59	<20	0.27	NEGATIVE	0.56	0.32	0.170	0.443
9011-05	09/23/97	146	<20	0.27	NEGATIVE	0.48	0.30	0.256	0.449
9011-06	09/25/97	133	<20	0.26	NEGATIVE	0.88	0.28	0.170	0.410
9011-07	09/30/97	87	<20	0.20	NEGATIVE	0.72	0.34	0.170	0.357
9011-08	10/02/97	403	<20	0.27	NEGATIVE	0.64	0.31	0.170	0.367
9011-09	10/07/97	8416	3324	0.27	NEGATIVE	1.04	0.33	0.170	0.370
9011-10	10/15/97	194100	55202	2.43	97.30	9.44	1.25	7.008	5.931
9011-11	10/17/97	436000	53577	4.05	106.10	13.44	2.02	12.649	9.475

PANEL MEMBER	**ORTHO VITROS ECi** ANTI-HIV ½ TEST DATE: 3/18/2015	**ABBOTT** PRISM ANTI-HIV ½ TEST DATE: 8/18/1998	**ORTHO ECi** ANTI-HIV TEST DATE: UNK	**BIOMERIEUX** ORGANON TEKNIKA ANTI-HIV-1 TEST DATE: 8/26/1998	**BIOMERIEUX** ORGANON VIRNOSTIKA HIV UNI-FORM II PLUS 0 TEST DATE: 9/2/1998	**BIOTEST** ANTI-HIV ½ TEST DATE: 9/24/1998	**DADE BEHRING** ENZYGNOST ANTI-HIV ½ PLUS TEST DATE: 8/31/1998	**GENETICS SYSTEMS** HIV ½ EIA TEST DATE: 8/25/1998
9011-01	0.43	0.23	0.24	0.58	0.37	0.007	0.02	0.112
9011-02	0.36	0.22	0.24	0.55	0.39	0.005	0.03	0.102
9011-03	0.38	0.19	0.22	0.48	0.38	0.026	0.00	0.105
9011-04	0.38	0.21	0.23	0.52	0.37	0.008	0.02	0.077
9011-05	0.33	0.22	0.22	0.53	0.36	0.000	0.01	0.115
9011-06	0.72	0.23	0.22	0.60	0.35	0.015	0.02	0.093
9011-07	0.36	0.21	0.22	0.58	0.37	0.020	0.01	0.080
9011-08	0.34	0.23	0.21	0.53	0.41	0.025	0.03	0.080
9011-09	0.36	0.22	0.20	0.57	0.38	0.013	0.02	0.089
9011-10	9.93	1.54	13.1	0.57	1.35	0.081	2.14	0.109
9011-11	18.7	2.60	22.0	0.57	1.94	0.548	4.79	0.109

Notes on Data:

1. This panel is for **Research Use Only. Not for use in diagnostic procedures.**
2. Data is generated at ZeptoMetrix™ LLC and by our Laboratory Partners in the United States and Europe.
3. All specimens collected in this longitudinal series are unadulterated 4% sodium citrated plasma samples collected from a single donor in the United States.
4. No preservatives have been added.
5. All material has been tested and found negative for HBsAg, Anti HCV by FDA cleared tests.

ZeptoMetrix LLC • 25 Kenwood Circle, Franklin, MA 02038 • Tel (508) 553-5800 • Fax (508) 520-1525
This Product was manufactured in a facility which has a Quality Management System that is ISO 13485 certified.
For Customer Support, please visit www.zeptometrix.com or email zepto.customerservice@antylia.com

PIHIV9011 Rev06
Effective Date: 10/25/2021

www.nrl-c.com

ZeptoMetrix®

HIV Seroconversion Panel Donor No. 64954
Part Number: HIV9011

Donor Demographics: Donor source plasma was collected from a U.S. licensed blood collection facility. Donor was un-treated and asymptomatic throughout the collection period.

PANEL MEMBER	**ORTHO** ANTI-HIV ½ ENHANCED ELISA TEST DATE: 9/15/1998	**CAMBRIDGE BIOTECH** HIV-1 WESTERN BLOT TEST DATE: UNK	**ABBOTT** PRISM HIV Ag/Ab COMBO 4TH Generation TEST DATE: UNK	**ABBOTT** MUREX HIV-1 Ab/Ag COMBO 4TH Generation TEST DATE: UNK	**ABBOTT** HIV Ag/Ab COMBO ARCHITECT 4TH Generation TEST DATE: UNK	**BIO-RAD** BioPlex 2200 HIV Ag-Ab 5th Generation TEST DATE: UNK	**BIO-RAD** BioPlex 2200 HIV-1 Ag 5th Generation TEST DATE: UNK	**BIO-RAD** BioPlex 2200 HIV-1 Ab 5th Generation TEST DATE: UNK
9011-01	0.04	NEGATIVE	0.09	0.280	0.09	0.11	0.11	0.07
9011-02	0.05	NEGATIVE	0.09	0.255	0.11	0.04	0.03	0.01
9011-03	0.04	NEGATIVE	0.07	0.265	0.07	0.08	0.06	0.02
9011-04	0.04	NEGATIVE	0.07	0.265	0.23	0.03	0.03	0.02
9011-05	0.05	NEGATIVE	0.08	0.275	0.16	0.10	0.10	0.04
9011-06	0.05	NEGATIVE	0.07	0.275	0.11	0.28	0.28	0.05
9011-07	0.04	NEGATIVE	0.09	0.245	0.11	0.08	0.08	0.03
9011-08	0.04	NEGATIVE	0.08	0.260	0.08	0.11	0.11	0.05
9011-09	0.04	P66(VF)	0.18	0.350	0.21	0.59	0.59	0.06
9011-10	5.47	P66(VF)	4.33	5.860	3.78	17.58	17.58	0.57
9011-11	8.50	P66(VF)	7.18	8.685	5.12	26.30	26.30	1.21

Notes on Data:

1. This panel is for **Research Use Only. Not for use in diagnostic procedures.**
2. Data is generated at ZeptoMetrix™ LLC and by our Laboratory Partners in the United States and Europe.
3. All specimens collected in this longitudinal series are unadulterated 4% sodium citrated plasma samples collected from a single donor in the United States.
4. No preservatives have been added.
5. All material has been tested and found negative for HBsAg, Anti HCV by FDA cleared tests.

ZeptoMetrix LLC • 25 Kenwood Circle, Franklin, MA 02038 • Tel (508) 553-5800 • Fax (508) 520-1525
This Product was manufactured in a facility which has a Quality Management System that is ISO 13485 certified.
For Customer Support, please visit www.zeptometrix.com or email zepto.customerservice@antylia.com

PIHIV9011 Rev06
Effective Date: 10/25/2021

www.h-h-c.com