HIV Seroconversion Panel Donor No. 66632
Part Number: HIV9034

Donor Demographics: Donor source plasma was collected from a U.S. licensed blood collection facility. Donor was un-treated and asymptomatic throughout the collection period.

PANEL MEMBER	DATE OF DRAW	**CHIRON** HIV-1 bDNA TEST DATE: 10/27/1999	**ABBOTT** HIV-1 p24 Ag TEST DATE: 8/29/2000	**ROCHE** ENZYMUM ANTI-HIV-1 COMBI TEST DATE: 11/15/1999	**COULTER** HIV-1 p24 Ag TEST DATE: 8/3/1999	**ABBOTT** AxSYM HIV AB ½ EIA TEST DATE: 11/17/1999	**ABBOTT** HIV AB ½ EIA rDNA TEST DATE: 5/11/1999	**ABBOTT** HIV Ab ½ EIA rDNA 3RD Generation TEST DATE: 11/15/1999	**ABBOTT** HIV Ab ½ EIA rDNA IMx TEST DATE: 6/7/2000
9034-01	02/01/98	<50	0.31	0.24	0.148	0.31	0.129	0.20	0.40
9034-02	02/03/98	<50	0.33	0.21	0.104	0.22	0.095	0.23	0.38
9034-03	02/08/98	<50	0.30	0.22	0.139	0.22	0.121	0.21	0.41
9034-04	02/10/98	<50	0.28	0.26	0.287	0.28	0.129	0.21	0.38
9034-05	02/24/98	<50	0.33	0.28	0.122	0.21	0.138	0.20	0.35
9034-06	02/26/98	<50	0.30	0.29	0.139	0.21	0.112	0.20	0.38
9034-07	03/05/98	<50	0.33	0.30	0.130	0.24	0.155	0.21	0.37
9034-08	03/07/98	<50	0.31	0.28	0.130	0.23	0.129	0.21	0.38
9034-09	03/12/98	<50	0.30	0.31	0.122	0.25	0.129	0.23	0.36
9034-10	03/14/98	54	0.33	0.29	0.130	0.22	0.147	0.25	0.43
9034-11	03/19/98	123310	0.87	0.34	2.191	0.69	0.147	0.23	0.47
9034-12	03/23/98	>500000	6.00	1.07	24.609	0.38	0.379	0.32	0.65
9034-13	03/26/98	>500000	11.08	2.44	29.391	0.88	0.974	1.07	1.43

PANEL MEMBER	**ABBOTT** MUREX HIV Ab ½ + O 3RD Generation TEST DATE: UNK	**ABBOTT** MUREX SUDS ANTI HIV-1 TEST DATE: 8/21/2000	**ABBOTT** PRISM ANTI-HIV ½ TEST DATE: 9/11/2000	**BIOMERIEUX** ORGANON VIRNOSTIKA HIV-1 MICRO ELISA SYSTEM TEST DATE: 8/3/1999	**BIOR-RAD** NOVAPATH HIV-1 IMMUNOBLOT TEST DATE: 9/4/2000	**DADE BEHRING** ENZYGNOST INTEGRAL HIV TEST DATE: 11/17/1999	**GENETICS SYSTEMS** HIV ½ PEPTIDE EIA TEST DATE: 8/4/1999	**ORTHO** ANTI-HIV ½ ENHANCED ELISA TEST DATE: 4/4/2000	**ORTHO VITROS ECi** ANTI-HIV ½ TEST DATE: 8/30/2018
9034-01	0.351	0.32	0.25	0.496	NEGATIVE	0.02	0.100	0.05	0.13
9034-02	0.306	0.27	0.23	0.504	NEGATIVE	0.03	0.118	0.05	0.13
9034-03	0.330	0.29	0.25	0.515	NEGATIVE	0.03	0.118	0.05	0.12
9034-04	0.281	0.27	0.27	0.515	NEGATIVE	0.02	0.118	0.05	0.12
9034-05	0.858	0.26	0.27	0.433	NEGATIVE	0.05	0.088	0.06	0.15
9034-06	0.285	0.30	0.25	0.470	NEGATIVE	0.02	0.206	0.07	0.14
9034-07	0.330	0.35	0.37	0.519	NEGATIVE	0.02	0.206	0.07	0.14
9034-08	0.389	0.38	0.33	0.444	NEGATIVE	0.02	0.088	0.05	0.13
9034-09	0.389	0.33	0.31	0.481	NEGATIVE	0.02	0.176	0.05	0.13
9034-10	0.337	0.29	0.32	0.456	NEGATIVE	0.03	0.176	0.05	0.13
9034-11	0.365	0.34	0.27	0.430	NEGATIVE	0.08	0.118	0.04	0.13
9034-12	0.420	0.39	0.41	0.537	NEGATIVE	0.75	0.206	0.19	0.18
9034-13	1.090	0.77	1.34	1.000	NEGATIVE	1.28	0.176	1.45	1.22

Notes on Data:

1. This panel is for **Research Use Only. Not for use in diagnostic procedures.**
2. Data is generated at ZeptoMetrix™ LLC and by our Laboratory Partners in the United States and Europe.
3. All specimens collected in this longitudinal series are unadulterated 4% sodium citrated plasma samples collected from a single donor in the United States.
4. No preservatives have been added.
5. All material has been tested and found negative for HBsAg, Anti HCV by FDA cleared tests.

ZeptoMetrix LLC • 25 Kenwood Circle, Franklin, MA 02038 • Tel (508) 553-5800 • Fax (508) 520-1525
This Product was manufactured in a facility which has a Quality Management System that is ISO 13485 certified.
For Customer Support, please visit www.zeptometrix.com or email zepto.customerservice@antylia.com

PIHIV9034 Rev06
Effective Date: 06/14/2021

www.hiv-c.com

ZeptoMetrix®

HIV Seroconversion Panel Donor No. 66632
Part Number: HIV9034

Donor Demographics: Donor source plasma was collected from a U.S. licensed blood collection facility. Donor was un-treated and asymptomatic throughout the collection period.

PANEL MEMBER	**ABBOTT** PRISM HIV Ag/Ab COMBO 4TH Generation TEST DATE: UNK	**ABBOTT** MUREX HIV-1 Ab/Ag COMBO 4TH Generation TEST DATE: UNK	**ABBOTT** HIV Ag/Ab COMBO ARCHITECT 4TH Generation TEST DATE: UNK	**BIO-RAD** BioPlex 2200 HIV Ag-Ab 5th Generation TEST DATE: UNK	**BIO-RAD** BioPlex 2200 HIV-1 Ag 5th Generation TEST DATE: UNK	**BIO-RAD** BioPlex 2200 HIV-1 Ab 5th Generation TEST DATE: UNK
9034-01	0.07	0.237	0.15	0.15	0.13	0.11
9034-02	0.12	0.227	0.13	0.08	0.08	0.05
9034-03	0.07	0.566	0.11	0.10	0.10	0.02
9034-04	0.07	0.217	0.10	0.14	0.14	0.04
9034-05	0.10	0.217	0.07	0.09	0.09	0.04
9034-06	0.07	0.222	0.12	0.07	0.07	0.01
9034-07	0.10	0.273	0.10	0.09	0.09	0.03
9034-08	0.10	0.258	0.15	0.13	0.13	0.06
9034-09	0.07	0.263	0.14	0.09	0.09	0.06
9034-10	0.07	0.253	0.10	0.12	0.12	0.08
9034-11	1.09	1.394	2.28	6.14	6.14	0.03
9034-12	17.64	14.601	29.75	84.03	84.03	0.05
9034-13	22.86	17.636	41.00	144.01	144.01	0.16

Notes on Data:

1. This panel is for **Research Use Only. Not for use in diagnostic procedures.**
2. Data is generated at ZeptoMetrix™ LLC and by our Laboratory Partners in the United States and Europe.
3. All specimens collected in this longitudinal series are unadulterated 4% sodium citrated plasma samples collected from a single donor in the United States.
4. No preservatives have been added.
5. All material has been tested and found negative for HBsAg, Anti HCV by FDA cleared tests.

ZeptoMetrix LLC • 25 Kenwood Circle, Franklin, MA 02038 • Tel (508) 553-5800 • Fax (508) 520-1525
This Product was manufactured in a facility which has a Quality Management System that is ISO 13485 certified.
For Customer Support, please visit www.zeptometrix.com or email zepto.customerservice@antylia.com

PIHIV9034 Rev06
Effective Date: 06/14/2021

www.h-h.com